Diagnostic Menu

DSD-201 Endoscope Reprocessor Diagnostic Functions			
0	Close all valves. All motors OFF.	26	Adjust LCD contrast
1	Activate detergent valve.	27	Adjust alarm volume
2	Activate water valve.	28	Lock cover
3	Activate alcohol pump.	29	Unlock cover
4	Activate air system (compressor and valve).	30	-- not used--
5	Activate air valve.	31	Adjust basin temp
6	Activate chamber valve	32	Adjust reservoir temp
7	Activate disinfectant pump.	33	-- not used--
8	Activate disinfect supply valve.	34	-- not used--
9	Activate disinfect overflow valve.	35	-- not used--
10	Activate disinfect return valve.	36	-- not used--
11	Activate drain valve.	37	-- not used--
12	Activate alcohol valve	38	-- not used--
13	Activate recirculation pump	39	-- not used--
14	Activate water inlet valve	40	Fluid flow sense inhibit
15	Activate disinfectant inlet valve	41	Cover sense inhibit
16	Activate recirculation inlet valve	42	Basin level sense inhibit
17	Activate recirculation chamber valve	43	Reservoir low level sense inhibit
18	Activate valves incrementally.	44	Reservoir high level sense inhibit
19	Activate detergent pump	45	Sheath sense inhibit
20	Activate station A LED's	46	Disinfectant warning inhibit
21	Activate station B LED's	47	Set alcohol sense enable
22	Activate system LED's	48	Set detergent sense enable
23	LCD test pattern	49	Set temperature monitor enable
24	Write counter to display	50	Turn leak tester off
25	Test keypad		

© 2001 *MediVators, Inc*